



HMP155A-L

Temperature and Relative Humidity Probe



Accurate, Wide Temp Range

Higher-end sensor where higher accuracy is required

Overview

The HMP155A provides reliable relative humidity (RH) and temperature measurements for a wide range of applications. It uses a HUMICAP® 180R capacitive thin film polymer sensor to

measure RH over the 0 to 100% RH range. A PRT measures temperature over the -80° to +60°C range. This rugged, accurate temperature/RH probe is manufactured by Vaisala.

Benefits and Features

- › Well-suited for long-term, unattended applications
- › Accurate and rugged
- › Mounts to a mast, crossarm, or user-supplied pole
- › Compatible with most Campbell Scientific dataloggers

Detailed Description

To reduce the current drain, power can be supplied to the HMP155A only during measurement when the sensor is connected to the datalogger's switched 12 V terminal.

Dataloggers that do not have a switched 12 V terminal, such as the CR510 or CR7, can use the SW12V switched 12 V device to switch power to the sensor only during measurement.

Specifications

Electromagnetic Compatibility	Complies with EMC standard EN61326-1 Electromagnetic
Filter Description	Sintered PTFE
Housing Material	PC
Housing Classification	IP66
Voltage Output	0 to 1 Vdc
Average Current Consumption	≤ 3 mA (analog output mode)

Operating Voltage	7 to 28 Vdc
Settling Time	2 s (at power up)
Tip Diameter	1.2 cm (0.5 in.)
Length	27.9 cm (11 in.)
Head Height	4 cm (1.6 in.)
Body Height	2.4 cm (0.9 in.)
Body Width	2.0 cm (0.8 in.)

Relative Humidity

Sensor	HUMICAP 180R
Measurement Range	0.8 to 100% RH (non-condensing)
Response Time	<ul style="list-style-type: none"> › The response time for the RH specification is for the HUMICAP 180R at 20°C in still air with sintered PTFE filter and a 0 to 75% RH step change. › 20 s (63% step change) › 60 s (90% step change)
Factory Calibration Uncertainty	<ul style="list-style-type: none"> › The factory calibration uncertainty is defined as ± 2 standard deviation limits. Uncertainty is at +20°C. Small variations are possible; see also the calibration certificate. › $\pm 0.6\%$ RH 0 to 40% RH › $\pm 1.0\%$ RH 40 to 97% RH

-NOTE-

Accuracy specifications include non-linearity, hysteresis, and repeatability.

Accuracy at 15° to 25°C	<ul style="list-style-type: none"> › $\pm 1\%$ RH (0 to 90% RH) › $\pm 1.7\%$ RH (90 to 100% RH)
Accuracy at -60° to -40°C	$\pm(1.4 + 0.032 \times \text{reading}) \% \text{ RH}$
Accuracy at -40° to -20°C	$\pm(1.2 + 0.012 \times \text{reading}) \% \text{ RH}$
Accuracy at -20° to +40°C	$\pm(1.0 + 0.008 \times \text{reading}) \% \text{ RH}$
Accuracy at 40° to 60°C	$\pm(1.2 + 0.012 \times \text{reading}) \% \text{ RH}$

Air Temperature

Sensor	Pt 100 RTD 1/3 class B IEC 751
Measurement Range	-80° to +60°C
Accuracy with Voltage Output	<ul style="list-style-type: none"> › $\pm(0.226 - 0.0028 \times \text{temperature})$ °C (-80° to +20°C) › $\pm(0.055 + 0.0057 \times \text{temperature})$ °C (+20° to +60°C)
Entire Temperature Range	Refer to graph in probe manual.

For comprehensive details, visit: www.campbellsci.com/hmp155a 



Campbell Scientific, Inc. | 815 W 1800 N | Logan, UT 84321-1784 | (435) 227-9120 | www.campbellsci.com
 AUSTRALIA | BRAZIL | CANADA | CHINA | COSTA RICA | FRANCE | GERMANY | THAILAND | SOUTH AFRICA | SPAIN | UK | USA

© 2018 Campbell Scientific, Inc. | 11/05/2018